

Analytical Report Nr.

AR-21-YL-000201-01

Sample code Nr.

560-2020-00010272

Date

11/01/2021

ANALYTICAL REPORT

Client Information

Eurofins Bactimm BV
Middenkampweg 19
Nijmegen the NETHERLANDS
+31 6 22477880
anjakroon@eurofins.com

For the attention of Mrs. Anja Kroon

Sample Information

Order Code: EUAA70-00009937
Reception Date: 28-Dec-2020
Analysis Starting Date: 28-Dec-2020
Analysis Ending Date: 11-Jan-2021
Sample code Nr. 560-2020-00010272
Sample described as: Masks

Requirements and decision rule

Customer requirements: EN 14683:2019+AC:2019 TYPE IIR
Decision Rule: Shared risk - Simple acceptance.

Information provided by the customer*

Client Reference: M800 E-H 38 gram splash BB
Sample Description:
Purchase Order Number:

Batch 2

Analytical Report Nr.

AR-21-YL-000201-01

Sample code Nr.

560-2020-00010272

Date

11/01/2021

SAMPLE PICTURE



Analytical Report Nr.

AR-21-YL-000201-01

Sample code Nr.

560-2020-00010272

Date

11/01/2021

CONCLUSION:

TEST PROPERTY	PASS	FAIL	REMARKS
• Microbial cleanliness (bioburden) EN ISO 11737-1:2018			
A-Mask	X		

Remark: Test has been performed as per application request

Analytical Report Nr.

AR-21-YL-000201-01

Sample code Nr.

560-2020-00010272

Date

11/01/2021

COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	A-Mask	Mask	Blue	---

Analytical Report Nr.

AR-21-YL-000201-01

Sample code Nr.

560-2020-00010272

Date

11/01/2021

MASKS TESTING

CAS No.

RESULTS

UNC.

LOQ

GUIDELINES

Analyses on:A-Mask• **Microbial cleanliness (bioburden)**

Analysis Ending Date: 11/01/2021

EN ISO 11737-1:2018

Bioburden

<30 cfu/g

-

≤ 30 cfu/g

✓ Pass

Complete test data reported at Annex.

Test covered by ACCREDIA accreditation scope n° 1827 L

Analytical Report Nr.

AR-21-YL-000201-01

Sample code Nr.

560-2020-00010272

Date

11/01/2021

Signed for and on behalf of Eurofins Textile Testing Spain:

Report electronically validated by

Axel Ferrando

Physical-Mechanical Lab Manager

EXPLANATORY NOTE

- ◆ Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- Test is subcontracted outside Eurofins group and is not accredited

N/A = Not Applicable

Eurofins Textile Testing Spain S.L.U is not responsible of the information supplied by the customer and reported as section "Information provided by the customer".

Eurofins General Sales Terms and Conditions Applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in section "Sample information" and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal, for those tests in which it is possible to evaluate the test uncertainty.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution provides a level of confidence of approximately 95%.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

If you happen to have any comments, please do it by sending email to textile_spain@eurofins.com and referring to this report number.

End Of Report



Cosmetics &
Personal Care



LAB N° 1827 L

Page: 1 of 1

TEST REPORT	Refer to Analytical Report Number			
SPONSOR	Eurofins Textile & Footwear Testing Spain			
	C/Germán Bernácer 4			
	03203 Elche (Alicante)			
	SPAIN			
TEST METHOD	Microbial cleanliness (Bioburden) – EN 14683:2019/AC 2019 par. 5.2.5 + App D			
TEST ITEM - INFORMATION FROM THE SPONSOR				
PRODUCT NAME	560-2020-00010272 - masks			
MATRIX OF THE PRODUCT	Face Mask			
BATCH	EUAA70-00009937	CODE	Not provided	
EUROFINS COSMETICS & PERSONAL CARE ITALY IDENTIFICATION				
MATERIAL ITEM ALIQUOT	N720AA1256-1			
PARCEL REGISTRATION N.	IP-N7-2020364-AAJ	RECEIVING DATE	29 Dec 2020	
ANALYSIS STARTING DATE	4 Jan 2021	ANALYSIS ENDING DATE	11 Jan 2021	
RESULTS	TOTAL BIOBURDEN	SPECIFICATION	RESULT	UNIT
	ALIQUOT 1	/	21.00	CFU/sample
		≤ 30	5.23	CFU/g
	ALIQUOT 2	/	12.00	CFU/sample
		≤ 30	3.13	CFU/g
	ALIQUOT 3	/	6.00	CFU/sample
		≤ 30	1.54	CFU/g
	ALIQUOT 4	/	6.00	CFU/sample
		≤ 30	1.56	CFU/g
	ALIQUOT 5	/	6.00	CFU/sample
		≤ 30	1.54	CFU/g

*This test report may not be reproduced in part unless expressly approved in writing by Eurofins Cosmetics & Personal Care Italy Srl.
The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor.
Information on the test item provided by the Sponsor are under Sponsor responsibility.*

Eurofins Cosmetics & Personal Care Italy Srl – via B.Buozzi 2, Vimodrone (Milano), Italy –
P.IVA / VAT Number: 05533561006
Tel: +39-022507151 – Fax: +39-0225071599 – E-mail: : InfoCosme@eurofins.com

Reviewed and electronically signed for Technical Supervisor Approval by
Martina Casini, Laboratory Manager
for Eurofins Cosmetic & Personal Care Italy Srl, on 11-Jan-2021 11:23:16 UTC+01:00

Operating requirements for surgical masks based on EN 14683: 2019+AC: 2019 standard

TEST	TYPE I	TYPE II	TYPE IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm ²)	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16
Microbial cleanliness (CFU/g)	≤ 30	≤ 30	≤ 30